

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042519\
 Data File : VN055245.D
 Acq On : 25 Apr 2019 14:44
 Operator : JC/SP
 Sample : K2203-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-03-042419-A

Quant Time: Apr 26 03:32:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	386374	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	683304	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	623521	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	195329	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	272922	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
35) Dibromofluoromethane	7.59	113	218954	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
50) Toluene-d8	10.09	98	839317	48.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.60%	
62) 4-Bromofluorobenzene	12.40	95	266015	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
Target Compounds						
16) Acetone	3.82	43	28309	16.375	ug/l	100
20) Methylene Chloride	4.55	84	153573	30.899	ug/l	98
43) Isopropyl Acetate	8.18	43	23233	2.803	ug/l #	80
95) Naphthalene	15.13	128	26203	7.280	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042519\
Data File : VN055245.D
Acq On : 25 Apr 2019 14:44
Operator : JC/SP
Sample : K2203-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-03-042419-A

Quant Time: Apr 26 03:32:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:52:53 2019
Response via : Initial Calibration

