

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042522\
 Data File : VN072181.D
 Acq On : 25 Apr 2022 15:51
 Operator : JC\MD
 Sample : PB144332TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144332TB

Quant Time: Apr 26 05:39:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	521611	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	891653	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	990353	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	375593	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	260976	43.826	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.660%
35) Dibromofluoromethane	8.022	113	253318	48.280	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.560%
50) Toluene-d8	10.439	98	1159219	54.008	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.020%
62) 4-Bromofluorobenzene	12.727	95	406881	58.614	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	117.220%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	15947	6.266	ug/l	95
20) Methylene Chloride	5.092	84	41661	7.767	ug/l	90
43) Isopropyl Acetate	8.557	43	44046	3.302	ug/l	93
95) Naphthalene	15.504	128	26782	5.250	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042522\
Data File : VN072181.D
Acq On : 25 Apr 2022 15:51
Operator : JC\MD
Sample : PB144332TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144332TB

Quant Time: Apr 26 05:39:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

