

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042522\
 Data File : VN072182.D
 Acq On : 25 Apr 2022 16:16
 Operator : JC\MD
 Sample : PB144332ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144332ZHE#01

Quant Time: Apr 26 05:39:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	508445	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	867083	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	937294	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	344584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	256088	44.119	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.240%
35) Dibromofluoromethane	8.016	113	241748	47.380	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.760%
50) Toluene-d8	10.439	98	1114635	53.403	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.800%
62) 4-Bromofluorobenzene	12.727	95	378432	56.061	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	112.120%
Target Compounds						
					Qvalue	
16) Acetone	4.298	43	15368	6.195	ug/l	93
20) Methylene Chloride	5.098	84	39961	7.643	ug/l	92
43) Isopropyl Acetate	8.557	43	44878	3.460	ug/l #	92
95) Naphthalene	15.498	128	7914	4.269	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042522\
Data File : VN072182.D
Acq On : 25 Apr 2022 16:16
Operator : JC\MD
Sample : PB144332ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144332ZHE#01

Quant Time: Apr 26 05:39:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

