

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042522\
 Data File : VN072184.D
 Acq On : 25 Apr 2022 17:04
 Operator : JC\MD
 Sample : PB144332ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144332ZHE#03

Quant Time: Apr 26 05:40:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	550609	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	934230	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1038009	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	398687	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	275200	43.781	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.560%
35) Dibromofluoromethane	8.016	113	263061	47.852	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.700%
50) Toluene-d8	10.439	98	1200166	53.368	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.740%
62) 4-Bromofluorobenzene	12.727	95	429949	59.114	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	118.220%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	15839	5.896	ug/l	90
20) Methylene Chloride	5.093	84	40905	7.224	ug/l	88
43) Isopropyl Acetate	8.557	43	64139	4.589	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042522\
Data File : VN072184.D
Acq On : 25 Apr 2022 17:04
Operator : JC\MD
Sample : PB144332ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144332ZHE#03

Quant Time: Apr 26 05:40:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

