

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042522\
 Data File : VN072185.D
 Acq On : 25 Apr 2022 17:28
 Operator : JC\MD
 Sample : PB144332ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB144332ZHE#04

Quant Time: Apr 26 05:40:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	560721	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	962560	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1066193	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	411314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	281471	43.971	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.940%
35) Dibromofluoromethane	8.016	113	271079	47.859	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.720%
50) Toluene-d8	10.439	98	1238937	53.470	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.940%
62) 4-Bromofluorobenzene	12.727	95	437651	58.402	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	116.800%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	14507	5.303	ug/l #	87
20) Methylene Chloride	5.104	84	41201	7.145	ug/l #	83
43) Isopropyl Acetate	8.557	43	65238	4.530	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042522\
Data File : VN072185.D
Acq On : 25 Apr 2022 17:28
Operator : JC\MD
Sample : PB144332ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB144332ZHE#04

Quant Time: Apr 26 05:40:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

