

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042522\
 Data File : VN072186.D
 Acq On : 25 Apr 2022 17:52
 Operator : JC\MD
 Sample : N2548-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-8S

Quant Time: Apr 26 05:40:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	516747	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	903348	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1023795	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	406431	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	265980	45.087	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.180%
35) Dibromofluoromethane	8.016	113	255985	48.157	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.320%
50) Toluene-d8	10.439	98	1165435	53.595	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.180%
62) 4-Bromofluorobenzene	12.727	95	427846	60.836	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	121.680%
Target Compounds						
					Qvalue	
16) Acetone	4.298	43	19008	7.539	ug/l	92
20) Methylene Chloride	5.098	84	44675	8.407	ug/l	91
43) Isopropyl Acetate	8.557	43	77983	5.770	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042522\
Data File : VN072186.D
Acq On : 25 Apr 2022 17:52
Operator : JC\MD
Sample : N2548-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-8S

Quant Time: Apr 26 05:40:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

