

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042522\
 Data File : VN072188.D
 Acq On : 25 Apr 2022 18:40
 Operator : JC\MD
 Sample : N2548-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-9

Quant Time: Apr 26 05:41:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	516148	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	891008	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	975403	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	365887	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	256070	43.457	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.920%
35) Dibromofluoromethane	8.022	113	247549	47.214	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.420%
50) Toluene-d8	10.439	98	1141935	53.242	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.480%
62) 4-Bromofluorobenzene	12.727	95	401054	57.816	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	115.640%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	30335	12.046	ug/l #	81
20) Methylene Chloride	5.093	84	35536	6.695	ug/l	89
43) Isopropyl Acetate	8.563	43	117857	8.842	ug/l #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042522\
Data File : VN072188.D
Acq On : 25 Apr 2022 18:40
Operator : JC\MD
Sample : N2548-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-9

Quant Time: Apr 26 05:41:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

