

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042523\
 Data File : VN077473.D
 Acq On : 25 Apr 2023 16:44
 Operator : JC\MD
 Sample : 02460-07 40X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-6

Quant Time: Apr 26 04:20:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	570666	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1075040	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	950714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	339530	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.589	65	372414	50.012	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.020%	
35) Dibromofluoromethane	8.172	113	326282	50.497	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.000%	
50) Toluene-d8	10.571	98	1295158	52.300	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.600%	
62) 4-Bromofluorobenzene	12.854	95	407760	50.327	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.660%	
Target Compounds						Qvalue
39) Methylcyclohexane	9.601	83	24471	1.990	ug/l	97
67) Ethyl Benzene	11.966	91	59060	1.668	ug/l	96
68) m/p-Xylenes	12.077	106	8458	0.614	ug/l	79
73) Isopropylbenzene	12.701	105	47593	1.511	ug/l	99
78) n-propylbenzene	13.036	91	132841	3.813	ug/l	99
84) 1,2,4-Trimethylbenzene	13.489	105	19797	0.906	ug/l	93
89) n-Butylbenzene	14.059	91	8885	0.478	ug/l #	76
95) Naphthalene	15.595	128	8447	2.061	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042523\
Data File : VN077473.D
Acq On : 25 Apr 2023 16:44
Operator : JC\MD
Sample : 02460-07 40X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-6

Quant Time: Apr 26 04:20:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

