

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042523\
 Data File : VN077475.D
 Acq On : 25 Apr 2023 17:32
 Operator : JC\MD
 Sample : 02460-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB-042023

Quant Time: Apr 26 04:21:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	588034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1091429	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	943390	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	320369	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	380397	49.575	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.160%
35) Dibromofluoromethane	8.172	113	334597	51.006	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.020%
50) Toluene-d8	10.572	98	1305094	51.910	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.820%
62) 4-Bromofluorobenzene	12.854	95	389926	47.403	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042523\
Data File : VN077475.D
Acq On : 25 Apr 2023 17:32
Operator : JC\MD
Sample : 02460-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB-042023

Quant Time: Apr 26 04:21:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

