

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042523\
 Data File : VN077477.D
 Acq On : 25 Apr 2023 18:21
 Operator : JC\MD
 Sample : 02460-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Quant Time: Apr 26 04:22:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	574468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1070198	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	942118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	331330	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	369486	49.290	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.580%
35) Dibromofluoromethane	8.172	113	331982	51.611	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.220%
50) Toluene-d8	10.572	98	1280997	51.962	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.920%
62) 4-Bromofluorobenzene	12.854	95	394222	48.876	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.760%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	17472	7.326	ug/l #	63
30) Chloroform	7.972	83	3952	0.303	ug/l	91
39) Methylcyclohexane	9.601	83	15177	1.240	ug/l #	87
73) Isopropylbenzene	12.701	105	13914	0.453	ug/l	96
78) n-propylbenzene	13.042	91	34458	1.014	ug/l	98
89) n-Butylbenzene	14.054	91	5943	0.328	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042523\
Data File : VN077477.D
Acq On : 25 Apr 2023 18:21
Operator : JC\MD
Sample : 02460-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Apr 26 04:22:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

