

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042524\
 Data File : VN081849.D
 Acq On : 25 Apr 2024 14:31
 Operator : JC\MD
 Sample : VN0425MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0425MBL01

Quant Time: Apr 26 05:34:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	247312	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	461677	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	399074	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	167745	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	194204	46.527	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.060%
35) Dibromofluoromethane	8.165	113	136301	49.035	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.060%
50) Toluene-d8	10.571	98	562058	51.484	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.960%
62) 4-Bromofluorobenzene	12.847	95	216173	50.091	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	100.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042524\
Data File : VN081849.D
Acq On : 25 Apr 2024 14:31
Operator : JC\MD
Sample : VN0425MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0425MBL01

Quant Time: Apr 26 05:34:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

