

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042524\
 Data File : VN081854.D
 Acq On : 25 Apr 2024 16:40
 Operator : JC\MD
 Sample : P2257-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41B

Quant Time: Apr 26 05:37:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	226201	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	416552	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	364870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	159869	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	176372	46.199	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.400%
35) Dibromofluoromethane	8.165	113	128143	51.094	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.180%
50) Toluene-d8	10.565	98	517790	52.567	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.140%
62) 4-Bromofluorobenzene	12.847	95	195871	50.304	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	100.600%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.512	59	10394	14.977	ug/l #	91
16) Acetone	4.424	43	25451	17.459	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042524\
Data File : VN081854.D
Acq On : 25 Apr 2024 16:40
Operator : JC\MD
Sample : P2257-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41B

Quant Time: Apr 26 05:37:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

