

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042524\
 Data File : VN081856.D
 Acq On : 25 Apr 2024 17:28
 Operator : JC\MD
 Sample : P2257-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 45A

Quant Time: Apr 26 05:38:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	222455	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	422572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	360394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	148303	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	175518	46.749	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.500%
35) Dibromofluoromethane	8.171	113	124663	48.998	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.000%
50) Toluene-d8	10.565	98	514910	51.530	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.060%
62) 4-Bromofluorobenzene	12.847	95	192475	48.728	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	97.460%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.524	59	9323	13.660	ug/l	# 92
16) Acetone	4.436	43	29383	21.468	ug/l	96
18) Methyl Acetate	5.024	43	12934	3.401	ug/l	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042524\
Data File : VN081856.D
Acq On : 25 Apr 2024 17:28
Operator : JC\MD
Sample : P2257-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
45A

Quant Time: Apr 26 05:38:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

