

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042619\
 Data File : VN055256.D
 Acq On : 25 Apr 2019 21:22
 Operator : JC/SP
 Sample : VN0426WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0426WBL01

Quant Time: Apr 26 05:30:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N042619W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Apr 26 05:19:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	66896	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	369364	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	300177	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	162270	30.73	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	102.43%
60) 4-Bromofluorobenzene	12.40	95	102644	22.28	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	74.27%
63) Toluene-d8	10.09	98	438484	30.75	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	102.50%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042619\
Data File : VN055256.D
Acq On : 25 Apr 2019 21:22
Operator : JC/SP
Sample : VN0426WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0426WBL01

Quant Time: Apr 26 05:30:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N042619W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Apr 26 05:19:18 2019
Response via : Initial Calibration

