

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN042619\
 Data File : VN055263.D
 Acq On : 26 Apr 2019 00:19
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Apr 26 05:51:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N042619W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Apr 26 05:19:18 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	100	0.00
2 M	Dichlorodifluoromethane	1.636	1.572	3.9	97	0.00
3 M	Chloromethane	2.288	2.249	1.7	101	0.00
4 M	Vinyl Chloride	1.865	1.902	-2.0	102	0.00
5 M	Bromomethane	1.039	1.072	-3.2	102	0.00
6 M	Chloroethane	1.039	1.059	-1.9	103	0.00
7 M	Trichlorofluoromethane	2.742	2.787	-1.6	103	0.00
8 T	Diethyl Ether	1.113	1.117	-0.4	105	0.00
9	1,1,2-Trichlorotrifluoroeth	1.745	1.681	3.7	97	0.00
10 M	1,1-Dichloroethene	1.625	1.580	2.8	98	0.00
11	Methyl Iodide	2.555	2.467	3.4	100	0.00
12	Methyl Acetate	2.358	2.277	3.4	96	0.00
13 M	Acrolein	0.105	0.093	11.4	96	0.00
14 M	Acrylonitrile	0.837	0.804	3.9	94	0.00
15 M	Acetone	0.202	0.172	14.9	81	0.00
16 M	Carbon Disulfide	4.415	4.179	5.3	100	0.00
17	Allyl chloride	3.498	3.367	3.7	102	0.00
18 M	Methylene Chloride	2.017	2.056	-1.9	101	0.00
19 M	trans-1,2-Dichloroethene	1.770	1.725	2.5	98	0.00
20 T	Diisopropyl ether	7.132	7.134	-0.0	99	0.00
21 M	1,1-Dichloroethane	3.706	3.769	-1.7	101	0.00
22 M	cis-1,2-Dichloroethene	2.050	2.052	-0.1	104	0.00
23 M	tert-Butyl Alcohol	0.171	0.152	11.1	94	0.00
24 M	Methyl tert-Butyl Ether	5.294	5.210	1.6	100	0.00
25 M	Chloroform	3.483	3.618	-3.9	104	0.00
26	Cyclohexane	3.219	3.140	2.5	101	0.00
27 s	1,2-Dichloroethane-d4	2.368	2.345	1.0	99	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
29	1,1-Dichloropropene	0.475	0.463	2.5	100	0.00
30 M	2-Butanone	0.193	0.171	11.4	87	0.00
31	2,2-Dichloropropane	0.450	0.368	18.2	82	0.00
32 M	1,1,1-Trichloroethane	0.529	0.519	1.9	100	0.00
33 M	Carbon Tetrachloride	0.441	0.431	2.3	102	0.00
34 M	Benzene	1.423	1.422	0.1	101	0.00
35	Methacrylonitrile	0.222	0.222	0.0	89	0.00
36 M	1,2-Dichloroethane	0.508	0.507	0.2	100	0.00
37 M	Trichloroethene	0.359	0.345	3.9	98	0.00
38	Methylcyclohexane	0.509	0.467	8.3	95	0.00
39 M	1,2-Dichloropropane	0.407	0.416	-2.2	102	0.00
40	Dibromomethane	0.236	0.237	-0.4	99	0.00
41 M	Bromodichloromethane	0.480	0.476	0.8	103	0.00
42 M	Vinyl Acetate	0.923	0.860	6.8	97	0.00
43	Ethyl Acetate	0.399	0.376	5.8	88	0.00
44	Isopropyl Acetate	0.735	0.675	8.2	91	0.00
45 T	1,4-Dioxane	0.004	0.004#	0.0	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN042619\
 Data File : VN055263.D
 Acq On : 26 Apr 2019 00:19
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Apr 26 05:51:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N042619W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Apr 26 05:19:18 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.346	0.307	11.3	89	0.00
47	n-amyl Acetate	0.523	0.409	21.8	85	0.00
48 M	t-1,3-Dichloropropene	0.472	0.413	12.5	94	0.00
49 T	cis-1,3-Dichloropropene	0.548	0.514	6.2	98	0.00
50 M	1,1,2-Trichloroethane	0.322	0.321	0.3	99	0.00
51	Ethyl methacrylate	0.438	0.388	11.4	96	0.00
52	1,3-Dichloropropane	0.570	0.539	5.4	96	0.00
53 M	Dibromochloromethane	0.333	0.304	8.7	96	0.00
54 M	1,2-Dibromoethane	0.312	0.297	4.8	99	0.00
55 M	2-Chloroethyl vinyl ether	0.132	0.113	14.4	94	0.00
56 M	Bromoform	0.218	0.192	11.9	92	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
58 M	4-Methyl-2-Pentanone	0.468	0.442	5.6	91	0.00
59 M	2-Hexanone	0.300	0.266	11.3	86	0.00
60 S	4-Bromofluorobenzene	0.460	0.450	2.2	99	0.00
61 M	Tetrachloroethene	0.389	0.412	-5.9	98	0.00
62 M	Toluene	1.697	1.731	-2.0	99	0.00
63 S	Toluene-d8	1.425	1.467	-2.9	100	0.00
64 M	Chlorobenzene	1.006	1.019	-1.3	100	0.00
65	1,1,1,2-Tetrachloroethane	0.375	0.382	-1.9	101	0.00
66 M	Ethyl Benzene	1.804	1.786	1.0	98	0.00
67 M	m/p-Xylenes	0.661	0.654	1.1	96	0.00
68 M	o-Xylene	0.659	0.664	-0.8	96	0.00
69 M	Styrene	1.021	0.990	3.0	98	0.00
70	Isopropylbenzene	1.736	1.749	-0.7	97	0.00
71 M	1,1,2,2-Tetrachloroethane	0.478	0.468	2.1	96	0.00
72	1,2,3-Trichloropropane	0.415	0.370	10.8	90	0.00
73	Bromobenzene	0.413	0.395	4.4	95	0.00
74	n-propylbenzene	1.937	1.837	5.2	93	0.00
75	2-Chlorotoluene	1.188	1.153	2.9	94	0.00
76	1,3,5-Trimethylbenzene	1.424	1.339	6.0	92	0.00
77	t-1,4-Dichloro-2-butene	0.110	0.088	20.0	90	0.00
78	4-Chlorotoluene	1.129	1.050	7.0	92	0.00
79	tert-butylbenzene	1.227	1.223	0.3	96	0.00
80	1,2,4-Trimethylbenzene	1.373	1.302	5.2	91	0.00
81	sec-Butylbenzene	1.622	1.547	4.6	93	0.00
82	p-Isopropyltoluene	1.345	1.214	9.7	89	0.00
83 M	1,3-Dichlorobenzene	0.669	0.609	9.0	93	0.00
84 M	1,4-Dichlorobenzene	0.632	0.553	12.5	90	0.00
85	n-Butylbenzene	1.026	0.805	21.5	85	0.00
86 T	Hexachloroethane	0.229	0.210	8.3	94	0.00
87 M	1,2-Dichlorobenzene	0.670	0.623	7.0	93	0.00
88	1,2-Dibromo-3-Chloropropane	0.070	0.059	15.7	89	0.00
89	1,2,4-Trichlorobenzene	0.265	0.178	32.8#	82	0.00
90	Hexachlorobutadiene	0.224	0.196	12.5	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN042619\
Data File : VN055263.D
Acq On : 26 Apr 2019 00:19
Operator : JC/SP
Sample : VSTDCCC020
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC020

Quant Time: Apr 26 05:51:55 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N042619W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Apr 26 05:19:18 2019
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	0.612	0.396	35.3#	84	0.00
92	1,2,3-Trichlorobenzene	0.276	0.191	30.8#	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0