

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042621\
 Data File : VN066755.D
 Acq On : 26 Apr 2021 12:22
 Operator : JC/MD
 Sample : VN0426MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0426MBL01

Quant Time: Apr 26 14:40:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.588	168	349089	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.516	114	536435	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	463347	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	187958	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.950	65	192537	45.15	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.30%
35) Dibromofluoromethane	7.511	113	165350	49.22	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.44%
50) Toluene-d8	10.029	98	639597	48.82	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.64%
62) 4-Bromofluorobenzene	12.346	95	190578	42.48	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.96%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042621\
Data File : VN066755.D
Acq On : 26 Apr 2021 12:22
Operator : JC/MD
Sample : VN0426MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0426MBL01

Quant Time: Apr 26 14:40:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

