

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042621\
 Data File : VN066756.D
 Acq On : 26 Apr 2021 12:57
 Operator : JC/MD
 Sample : VN0426WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0426WBL01

Quant Time: Apr 26 14:40:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.588	168	339432	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.516	114	515779	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	450226	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	197607	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	187071	45.11	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	90.22%
35) Dibromofluoromethane	7.510	113	159721	49.45	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.90%
50) Toluene-d8	10.029	98	620499	49.26	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.52%
62) 4-Bromofluorobenzene	12.346	95	188902	43.79	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.58%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042621\
Data File : VN066756.D
Acq On : 26 Apr 2021 12:57
Operator : JC/MD
Sample : VN0426WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0426WBL01

Quant Time: Apr 26 14:40:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

