

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042621\
 Data File : VN066764.D
 Acq On : 26 Apr 2021 16:16
 Operator : JC/MD
 Sample : PB135974TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135974TB

Quant Time: Apr 27 06:13:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	377870	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	578150	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	527466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	235197	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	212635	46.06	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	92.12%		
35) Dibromofluoromethane	7.513	113	176077	48.63	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.26%		
50) Toluene-d8	10.029	98	713987	50.57	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.14%		
62) 4-Bromofluorobenzene	12.346	95	242739	50.20	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	100.40%		
Target Compounds						
16) Acetone	3.759	43	39095	22.62	ug/l	96
20) Methylene Chloride	4.472	84	21607	4.21	ug/l	99
43) Isopropyl Acetate	8.095	43	36766	4.08	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042621\
Data File : VN066764.D
Acq On : 26 Apr 2021 16:16
Operator : JC/MD
Sample : PB135974TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135974TB

Quant Time: Apr 27 06:13:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

