

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042621\
 Data File : VN066765.D
 Acq On : 26 Apr 2021 16:40
 Operator : JC/MD
 Sample : PB135974ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135974ZHE#01

Quant Time: Apr 27 06:13:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	394380	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	616186	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	557395	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	245738	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	230106	47.76	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.52%		
35) Dibromofluoromethane	7.513	113	188067	48.74	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.48%		
50) Toluene-d8	10.029	98	755878	50.23	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	100.46%		
62) 4-Bromofluorobenzene	12.346	95	258286	50.12	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	100.24%		
Target Compounds						
16) Acetone	3.758	43	37323	20.69	ug/l	97
20) Methylene Chloride	4.472	84	25293	4.72	ug/l	99
43) Isopropyl Acetate	8.093	43	38595	4.02	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042621\
Data File : VN066765.D
Acq On : 26 Apr 2021 16:40
Operator : JC/MD
Sample : PB135974ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135974ZHE#01

Quant Time: Apr 27 06:13:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

