

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042621\
 Data File : VN066754.D
 Acq On : 26 Apr 2021 11:50
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 26 14:39:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	142	0.00
2 T	Dichlorodifluoromethane	50.000	44.595	10.8	124	0.00
3 P	Chloromethane	50.000	41.768	16.5	122	0.00
4 C	Vinyl Chloride	50.000	47.039	5.9#	128	0.00
5 T	Bromomethane	50.000	51.117	-2.2	134	0.02
6 T	Chloroethane	50.000	46.611	6.8	127	0.00
7 T	Trichlorofluoromethane	50.000	47.958	4.1	133	0.00
8 T	Diethyl Ether	50.000	46.564	6.9	128	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.568	4.9	133	0.00
10 T	Methyl Iodide	50.000	45.072	9.9	124	0.00
11 T	Tert butyl alcohol	250.000	245.981	1.6	137	-0.01
12 CM	1,1-Dichloroethene	50.000	46.155	7.7#	127	0.00
13 T	Acrolein	250.000	138.125	44.8#	79	0.00
14 T	Allyl chloride	50.000	43.875	12.3	118	0.00
15 T	Acrylonitrile	250.000	231.907	7.2	124	0.00
16 T	Acetone	250.000	263.412	-5.4	157	0.00
17 T	Carbon Disulfide	50.000	43.584	12.8	122	0.00
18 T	Methyl Acetate	50.000	43.342	13.3	125	0.00
19 T	Methyl tert-butyl Ether	50.000	46.985	6.0	125	0.00
20 T	Methylene Chloride	50.000	40.343	19.3	127	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.685	6.6	127	0.00
22 T	Diisopropyl ether	50.000	43.715	12.6	117	0.00
23 T	Vinyl Acetate	250.000	222.599	11.0	116	0.00
24 P	1,1-Dichloroethane	50.000	44.482	11.0	124	0.00
25 T	2-Butanone	250.000	240.379	3.8	128	0.00
26 T	2,2-Dichloropropane	50.000	45.661	8.7	128	0.00
27 T	cis-1,2-Dichloroethene	50.000	46.081	7.8	125	0.00
28 T	Bromochloromethane	50.000	48.520	3.0	136	0.00
29 T	Tetrahydrofuran	250.000	224.036	10.4	120	0.00
30 C	Chloroform	50.000	45.846	8.3#	125	0.00
31 T	Cyclohexane	50.000	43.371	13.3	125	0.00
32 T	1,1,1-Trichloroethane	50.000	46.075	7.8	125	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.342	5.3	138	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	137	0.00
35 S	Dibromofluoromethane	50.000	52.164	-4.3	144	0.00
36 T	1,1-Dichloropropene	50.000	48.605	2.8	127	0.00
37 T	Ethyl Acetate	50.000	43.893	12.2	118	0.00
38 T	Carbon Tetrachloride	50.000	48.391	3.2	128	0.00
39 T	Methylcyclohexane	50.000	51.947	-3.9	133	0.00
40 TM	Benzene	50.000	47.793	4.4	124	0.00
41 T	Methacrylonitrile	50.000	51.139	-2.3	153	0.01
42 TM	1,2-Dichloroethane	50.000	45.699	8.6	121	0.00
43 T	Isopropyl Acetate	50.000	45.407	9.2	118	0.00
44 TM	Trichloroethene	50.000	50.729	-1.5	132	0.00
45 C	1,2-Dichloropropane	50.000	46.928	6.1#	124	0.00
46 T	Dibromomethane	50.000	48.612	2.8	125	0.00
47 T	Bromodichloromethane	50.000	48.051	3.9	126	0.00
48 T	Methyl methacrylate	50.000	47.786	4.4	118	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042621\
 Data File : VN066754.D
 Acq On : 26 Apr 2021 11:50
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 26 14:39:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 22 08:25:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1230.554	-23.1	152	0.00
50 S	Toluene-d8	50.000	53.877	-7.8	149	0.00
51 T	4-Methyl-2-Pentanone	250.000	236.839	5.3	119	0.00
52 CM	Toluene	50.000	50.471	-0.9#	128	0.00
53 T	t-1,3-Dichloropropene	50.000	50.817	-1.6	127	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.953	0.1	127	0.00
55 T	1,1,2-Trichloroethane	50.000	48.813	2.4	127	0.00
56 T	Ethyl methacrylate	50.000	44.536	10.9	126	0.00
57 T	1,3-Dichloropropane	50.000	48.798	2.4	125	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	232.843	6.9	141	0.00
59 T	2-Hexanone	250.000	219.272	12.3	124	0.00
60 T	Dibromochloromethane	50.000	50.898	-1.8	129	0.00
61 T	1,2-Dibromoethane	50.000	50.627	-1.3	127	0.00
62 S	4-Bromofluorobenzene	50.000	54.972	-9.9	145	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	135	0.00
64 T	Tetrachloroethene	50.000	49.242	1.5	132	0.00
65 PM	Chlorobenzene	50.000	50.027	-0.1	129	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.289	-2.6	130	0.00
67 C	Ethyl Benzene	50.000	51.945	-3.9#	127	0.00
68 T	m/p-Xylenes	100.000	106.807	-6.8	129	0.00
69 T	o-Xylene	50.000	51.911	-3.8	128	0.00
70 T	Styrene	50.000	45.202	9.6	126	0.00
71 P	Bromoform	50.000	52.327	-4.7	127	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	131	0.00
73 T	Isopropylbenzene	50.000	50.941	-1.9	128	0.00
74 T	N-ethyl acetate	50.000	53.405	-6.8	124	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.046	5.9	126	0.00
76 T	1,2,3-Trichloropropane	50.000	45.883	8.2	116	0.00
77 T	Bromobenzene	50.000	49.756	0.5	127	0.00
78 T	n-propylbenzene	50.000	52.905	-5.8	127	0.00
79 T	2-Chlorotoluene	50.000	46.251	7.5	124	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.989	-2.0	127	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.999	8.0	128	0.00
82 T	4-Chlorotoluene	50.000	51.814	-3.6	126	0.00
83 T	tert-Butylbenzene	50.000	48.571	2.9	130	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.787	-3.6	127	0.00
85 T	sec-Butylbenzene	50.000	51.858	-3.7	128	0.00
86 T	p-Isopropyltoluene	50.000	55.673	-11.3	132	0.00
87 T	1,3-Dichlorobenzene	50.000	51.413	-2.8	128	0.00
88 T	1,4-Dichlorobenzene	50.000	48.711	2.6	130	0.00
89 T	n-Butylbenzene	50.000	51.804	-3.6	134	0.00
90 T	Hexachloroethane	50.000	46.809	6.4	125	0.00
91 T	1,2-Dichlorobenzene	50.000	49.324	1.4	128	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	41.273	17.5	124	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.419	-0.8	137	0.00
94 T	Hexachlorobutadiene	50.000	51.200	-2.4	140	0.00
95 T	Naphthalene	50.000	48.628	2.7	133	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042621\
Data File : VN066754.D
Acq On : 26 Apr 2021 11:50
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 26 14:39:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042121W.M
Quant Title : SW846 8260
QLast Update : Thu Apr 22 08:25:55 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.532	0.9	134	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6