


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042622\
Data File : VN072195.D
Acq On : 26 Apr 2022 12:49
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Title : SW846 8260
Last Update : Tue Apr 26 18:50:11 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1817

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	9921	PASS
75	95	30	60	54.7	27837	PASS
95	95	100	100	100.0	50883	PASS
96	95	5	9	6.8	3460	PASS
173	174	0.00	2	1.2	456	PASS
174	95	50	100	72.9	37075	PASS
175	174	5	9	7.6	2813	PASS
176	174	95	101	98.6	36549	PASS
177	176	5	9	6.2	2266	PASS