

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042624\
 Data File : VN081872.D
 Acq On : 26 Apr 2024 12:47
 Operator : JC\MD
 Sample : VN0426WBS02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0426WBS02

Quant Time: Apr 26 22:58:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	246880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	452375	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	377597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	161162	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	188645	45.275	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.540%
35) Dibromofluoromethane	8.171	113	134860	49.514	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.020%
50) Toluene-d8	10.565	98	548184	51.246	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.500%
62) 4-Bromofluorobenzene	12.847	95	193985	45.874	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	91.740%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	58962	17.014	ug/l	100
3) Chloromethane	2.359	50	57467	15.895	ug/l	100
4) Vinyl Chloride	2.518	62	55804	17.051	ug/l	99
5) Bromomethane	2.965	94	27341	16.169	ug/l	98
6) Chloroethane	3.130	64	29749	17.985	ug/l	91
7) Trichlorofluoromethane	3.500	101	88497	17.481	ug/l	96
8) Diethyl Ether	3.965	74	38651	17.973	ug/l	87
9) 1,1,2-Trichlorotrifluo...	4.383	101	55364	18.836	ug/l	93
10) Methyl Iodide	4.594	142	48839	18.971	ug/l #	90
11) Tert butyl alcohol	5.524	59	55667	73.491	ug/l	96
12) 1,1-Dichloroethene	4.341	96	54895	18.359	ug/l	86
13) Acrolein	4.177	56	59632	96.421	ug/l	98
14) Allyl chloride	5.024	41	105891	16.742	ug/l #	88
15) Acrylonitrile	5.724	53	141262	77.665	ug/l	99
16) Acetone	4.436	43	106006	82.369	ug/l	99
17) Carbon Disulfide	4.712	76	147926	16.459	ug/l #	94
18) Methyl Acetate	5.024	43	70881	16.795	ug/l	90
19) Methyl tert-butyl Ether	5.800	73	195472	17.324	ug/l	96
20) Methylene Chloride	5.277	84	58578	16.405	ug/l #	71
21) trans-1,2-Dichloroethene	5.783	96	55524	17.031	ug/l #	79
22) Diisopropyl ether	6.665	45	207018	16.915	ug/l #	93
23) Vinyl Acetate	6.600	43	847412	85.214	ug/l	95
24) 1,1-Dichloroethane	6.571	63	110790	17.392	ug/l	97
25) 2-Butanone	7.482	43	196046	82.898	ug/l #	85
26) 2,2-Dichloropropane	7.488	77	103161	17.604	ug/l	99
27) cis-1,2-Dichloroethene	7.488	96	67975	17.843	ug/l	87
28) Bromochloromethane	7.818	49	56151	18.546	ug/l	83
29) Tetrahydrofuran	7.835	42	133631	78.797	ug/l	89
30) Chloroform	7.965	83	108533	17.188	ug/l	95
31) Cyclohexane	8.259	56	107201	17.079	ug/l	86
32) 1,1,1-Trichloroethane	8.165	97	92138	16.362	ug/l	95
36) 1,1-Dichloropropene	8.371	75	80626	18.195	ug/l	97
37) Ethyl Acetate	7.565	43	82739	17.270	ug/l	96
38) Carbon Tetrachloride	8.359	117	75024	17.324	ug/l	98
39) Methylcyclohexane	9.600	83	98453	18.295	ug/l	91
40) Benzene	8.606	78	247146	18.580	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042624\
 Data File : VN081872.D
 Acq On : 26 Apr 2024 12:47
 Operator : JC\MD
 Sample : VN0426WBS02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0426WBS02

Quant Time: Apr 26 22:58:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.782	41	49092	16.059	ug/l	89
42) 1,2-Dichloroethane	8.665	62	80902	16.379	ug/l	99
43) Isopropyl Acetate	8.688	43	151459	16.813	ug/l	95
44) Trichloroethene	9.353	130	57273	18.675	ug/l	92
45) 1,2-Dichloropropane	9.623	63	63647	18.283	ug/l	98
46) Dibromomethane	9.712	93	41469	17.235	ug/l	92
47) Bromodichloromethane	9.888	83	84750	16.701	ug/l	99
48) Methyl methacrylate	9.682	41	69783	16.227	ug/l	89
49) 1,4-Dioxane	9.700	88	20319	310.864	ug/l #	73
51) 4-Methyl-2-Pentanone	10.447	43	426569	84.116	ug/l	94
52) Toluene	10.629	92	149832	18.809	ug/l	99
53) t-1,3-Dichloropropene	10.835	75	95343	17.268	ug/l	98
54) cis-1,3-Dichloropropene	10.312	75	104522	18.367	ug/l #	87
55) 1,1,2-Trichloroethane	11.018	97	54541	17.409	ug/l	97
56) Ethyl methacrylate	10.876	69	100519	16.850	ug/l #	82
57) 1,3-Dichloropropane	11.165	76	99989	17.952	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.159	63	225270	82.892	ug/l	92
59) 2-Hexanone	11.194	43	323606	85.746	ug/l	95
60) Dibromochloromethane	11.359	129	58869	17.741	ug/l	99
61) 1,2-Dibromoethane	11.470	107	53513	17.862	ug/l	99
64) Tetrachloroethene	11.106	164	53927	19.664	ug/l	96
65) Chlorobenzene	11.894	112	152308	18.471	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.959	131	48500	18.098	ug/l	93
67) Ethyl Benzene	11.964	91	281068	18.328	ug/l	97
68) m/p-Xylenes	12.070	106	206985	37.840	ug/l	90
69) o-Xylene	12.400	106	105189	19.697	ug/l	96
70) Styrene	12.412	104	174199	18.856	ug/l	94
71) Bromoform	12.576	173	34076	16.174	ug/l #	98
73) Isopropylbenzene	12.694	105	259227	18.091	ug/l	99
74) N-amyl acetate	12.494	43	130375	16.343	ug/l	91
75) 1,1,2,2-Tetrachloroethane	12.941	83	69306	16.107	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	103523	24.775	ug/l	65
77) Bromobenzene	12.982	156	55583	18.537	ug/l	80
78) n-propylbenzene	13.035	91	310249	17.825	ug/l	99
79) 2-Chlorotoluene	13.123	91	189686	17.130	ug/l	92
80) 1,3,5-Trimethylbenzene	13.170	105	212321	17.511	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.735	75	35494	17.406	ug/l	88
82) 4-Chlorotoluene	13.223	91	186280	17.327	ug/l	92
83) tert-Butylbenzene	13.435	119	181360	17.961	ug/l	94
84) 1,2,4-Trimethylbenzene	13.482	105	215569	17.803	ug/l	96
85) sec-Butylbenzene	13.617	105	263768	18.545	ug/l	99
86) p-Isopropyltoluene	13.729	119	211808	18.546	ug/l	97
87) 1,3-Dichlorobenzene	13.729	146	100436	17.952	ug/l	98
88) 1,4-Dichlorobenzene	13.811	146	100791	18.294	ug/l	94
89) n-Butylbenzene	14.058	91	189868	17.861	ug/l	97
90) Hexachloroethane	14.335	117	40243	17.913	ug/l	94
91) 1,2-Dichlorobenzene	14.106	146	99211	18.498	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.717	75	17076	16.300	ug/l	80
93) 1,2,4-Trichlorobenzene	15.394	180	57050	19.013	ug/l	97
94) Hexachlorobutadiene	15.505	225	21661	19.586	ug/l	98
95) Naphthalene	15.641	128	200536	17.687	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	56779	19.709	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042624\
Data File : VN081872.D
Acq On : 26 Apr 2024 12:47
Operator : JC\MD
Sample : VN0426WBS02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02

Quant Time: Apr 26 22:58:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042624\
 Data File : VN081872.D
 Acq On : 26 Apr 2024 12:47
 Operator : JC\MD
 Sample : VN0426WBS02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0426WBS02

Quant Time: Apr 26 22:58:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.224 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :





VN0426WBS02

Tgt Ion:168 Resp: 246880

Ion Ratio Lower Upper

168 100

99 64.3 48.2 72.4



#2

Dichlorodifluoromethane

Concen: 17.014 ug/l

RT: 2.124 min Scan# 29

Delta R.T. -0.000 min

Lab File: VN081872.D




Acq: 26 Apr 2024 12:47


Tgt Ion: 85 Resp: 58962

Ion Ratio Lower Upper

85 100

87 31.6 15.8 47.4









#3
Chloromethane
Concen: 15.895 ug/l
RT: 2.359 min Scan# 69
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 50 Resp: 57467
Ion Ratio Lower Upper
50 100
52 31.8 25.5 38.3



#4
Vinyl Chloride
Concen: 17.051 ug/l
RT: 2.518 min Scan# 96
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 62 Resp: 55804
Ion Ratio Lower Upper
62 100
64 30.1 24.6 37.0









#5
Bromomethane
Concen: 16.169 ug/l
RT: 2.965 min Scan# 11
Delta R.T. 0.017 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 94 Resp: 27341
Ion Ratio Lower Upper
94 100
96 92.0 74.8 112.2



#6
Chloroethane
Concen: 17.985 ug/l
RT: 3.130 min Scan# 200
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 64 Resp: 29749
Ion Ratio Lower Upper
64 100
66 35.8 24.7 37.1









#7
Trichlorofluoromethane
Concen: 17.481 ug/l
RT: 3.500 min Scan# 20
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion:101 Resp: 88497
Ion Ratio Lower Upper
101 100
103 66.6 50.6 75.8



#8
Diethyl Ether
Concen: 17.973 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.012 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 74 Resp: 38651
Ion Ratio Lower Upper
74 100
45 97.0 42.4 127.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.836 ug/l
RT: 4.383 min Scan# 412
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion:101 Resp: 55364
Ion Ratio Lower Upper
101 100
85 44.6 37.3 55.9
151 66.2 59.2 88.8



#10
Methyl Iodide
Concen: 18.971 ug/l
RT: 4.594 min Scan# 449
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion:142 Resp: 48839
Ion Ratio Lower Upper
142 100
127 37.8 35.8 53.6
141 11.2 11.4 17.2#









#11
Tert butyl alcohol
Concen: 73.491 ug/l
RT: 5.524 min Scan# 606
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 59 Resp: 55667
Ion Ratio Lower Upper
59 100
57 12.5 8.9 13.3



#12
1,1-Dichloroethene
Concen: 18.359 ug/l
RT: 4.341 min Scan# 406
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 96 Resp: 54895
Ion Ratio Lower Upper
96 100
61 169.6 119.7 179.5
98 55.6 49.6 74.4





#13

Acrolein

Concen: 96.421 ug/l

RT: 4.177 min Scan# 31

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :





VN0426WBS02

Tgt Ion: 56 Resp: 59632

Ion Ratio Lower Upper

56 100

55 71.7 55.8 83.8



#14

Allyl chloride

Concen: 16.742 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47




Tgt Ion: 41 Resp: 105891


Ion Ratio Lower Upper

41 100

39 72.8 65.4 98.0

76 35.5 37.3 55.9#









#15
Acrylonitrile
Concen: 77.665 ug/l
RT: 5.724 min Scan# 641
Delta R.T. 0.012 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 53 Resp: 141262
Ion Ratio Lower Upper
53 100
52 83.7 66.1 99.1
51 37.4 29.4 44.2



#16
Acetone
Concen: 82.369 ug/l
RT: 4.436 min Scan# 422
Delta R.T. 0.012 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 43 Resp: 106006
Ion Ratio Lower Upper
43 100
58 31.0 25.2 37.8









#17
Carbon Disulfide
Concen: 16.459 ug/l
RT: 4.712 min Scan# 469
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 76 Resp: 147926
Ion Ratio Lower Upper
76 100
78 7.4 7.5 11.3#



#18
Methyl Acetate
Concen: 16.795 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 43 Resp: 70881
Ion Ratio Lower Upper
43 100
74 22.3 22.2 33.2









#19
Methyl tert-butyl Ether
Concen: 17.324 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 73 Resp: 195472
Ion Ratio Lower Upper
73 100
57 22.3 16.5 24.7



#20
Methylene Chloride
Concen: 16.405 ug/l
RT: 5.277 min Scan# 565
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 84 Resp: 58578
Ion Ratio Lower Upper
84 100
49 152.1 84.9 127.3#
51 47.6 27.0 40.6#
86 71.2 52.3 78.5









#21
trans-1,2-Dichloroethene
Concen: 17.031 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 96 Resp: 55524
Ion Ratio Lower Upper
96 100
61 152.7 96.5 144.7#
98 58.4 51.3 76.9



#22
Diisopropyl ether
Concen: 16.915 ug/l
RT: 6.665 min Scan# 801
Delta R.T. -0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 45 Resp: 207018
Ion Ratio Lower Upper
45 100
43 57.4 44.1 66.1
87 25.3 27.0 40.6#
59 11.3 9.6 14.4









#23
 Vinyl Acetate
 Concen: 85.214 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. -0.000 min
 Lab File: VN081872.D
 Acq: 26 Apr 2024 12:47




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0426WBS02


Tgt Ion: 43 Resp: 847412
 Ion Ratio Lower Upper
 43 100
 86 10.9 10.2 15.4



#24
 1,1-Dichloroethane
 Concen: 17.392 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. 0.006 min
 Lab File: VN081872.D
 Acq: 26 Apr 2024 12:47

Tgt Ion: 63 Resp: 110790
 Ion Ratio Lower Upper
 63 100
 98 7.4 3.6 10.9
 100 3.3 2.6 7.8









#25
2-Butanone
Concen: 82.898 ug/l
RT: 7.482 min Scan# 940
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 43 Resp: 196046
Ion Ratio Lower Upper
43 100
72 23.3 25.3 37.9#



#26
2,2-Dichloropropane
Concen: 17.604 ug/l
RT: 7.488 min Scan# 941
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 77 Resp: 103161
Ion Ratio Lower Upper
77 100
97 22.0 11.3 33.9









#27
cis-1,2-Dichloroethene
Concen: 17.843 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 96 Resp: 67975
Ion Ratio Lower Upper
96 100
61 154.9 0.0 268.0
98 65.8 0.0 126.8



#28
Bromochloromethane
Concen: 18.546 ug/l
RT: 7.818 min Scan# 997
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 49 Resp: 56151
Ion Ratio Lower Upper
49 100
129 1.9 0.0 4.6
130 68.0 66.7 100.1





#29

Tetrahydrofuran

Concen: 78.797 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :

VN0426WBS02





Tgt Ion: 42 Resp: 133631

Ion Ratio Lower Upper

42 100

72 41.5 39.3 58.9

71 39.5 37.4 56.0



#30

Chloroform

Concen: 17.188 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.000 min

Lab File: VN081872.D




Acq: 26 Apr 2024 12:47


Tgt Ion: 83 Resp: 108533

Ion Ratio Lower Upper

83 100

85 59.6 50.7 76.1





#31

Cyclohexane

Concen: 17.079 ug/l

RT: 8.259 min Scan# 1072

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :

VN0426WBS02





Tgt Ion: 56 Resp: 107201

Ion Ratio Lower Upper

56 100

69 29.6 24.4 36.6

84 79.2 77.0 115.4



#32

1,1,1-Trichloroethane

Concen: 16.362 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47




Tgt Ion: 97 Resp: 92138


Ion Ratio Lower Upper

97 100

99 60.8 50.3 75.5

61 51.0 36.6 55.0





#33

1,2-Dichloroethane-d4

Concen: 45.275 ug/l

RT: 8.582 min Scan# 1126

Delta R.T. 0.006 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :





VN0426WBS02

Tgt Ion: 65 Resp: 188645

Ion Ratio Lower Upper

65 100

67 51.2 0.0 103.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47




Tgt Ion: 114 Resp: 452375


Ion Ratio Lower Upper

114 100

63 22.9 0.0 41.0

88 16.5 0.0 31.4









#35
Dibromofluoromethane
Concen: 49.514 ug/l
RT: 8.171 min Scan# 1056
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion:113 Resp: 134860
Ion Ratio Lower Upper
113 100
111 103.7 83.4 125.2
192 17.1 14.0 21.0



#36
1,1-Dichloropropene
Concen: 18.195 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 75 Resp: 80626
Ion Ratio Lower Upper
75 100
110 32.2 17.8 53.5
77 31.2 25.0 37.4









#37
Ethyl Acetate
Concen: 17.270 ug/l
RT: 7.565 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 43 Resp: 82739
Ion Ratio Lower Upper
43 100
61 12.7 11.3 16.9
70 12.6 8.7 13.1



#38
Carbon Tetrachloride
Concen: 17.324 ug/l
RT: 8.359 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 117 Resp: 75024
Ion Ratio Lower Upper
117 100
119 96.5 76.5 114.7
121 25.9 23.9 35.9





#39

Methylcyclohexane

Concen: 18.295 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. -0.000 min

Lab File: VN081872.D


Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :

VN0426WBS02






Tgt Ion: 83 Resp: 98453

Ion Ratio Lower Upper

83 100

55 77.1 53.0 79.6

98 45.2 37.1 55.7



#40

Benzene

Concen: 18.580 ug/l

RT: 8.606 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VN081872.D




Acq: 26 Apr 2024 12:47


Tgt Ion: 78 Resp: 247146

Ion Ratio Lower Upper

78 100

77 24.4 20.0 30.0









#41
Methacrylonitrile
Concen: 16.059 ug/l
RT: 7.782 min Scan# 991
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion:	41	Resp:	49092
Ion	Ratio	Lower	Upper
41	100		
39	61.1	46.6	69.8
67	68.7	67.3	100.9
52	30.5	27.3	40.9



#42
1,2-Dichloroethane
Concen: 16.379 ug/l
RT: 8.665 min Scan# 1141
Delta R.T. -0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion:	62	Resp:	80902
Ion	Ratio	Lower	Upper
62	100		
98	10.3	0.0	19.6









#43
Isopropyl Acetate
Concen: 16.813 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 43 Resp: 151459
Ion Ratio Lower Upper
43 100
61 25.9 22.8 34.2
87 12.9 12.0 18.0



#44
Trichloroethene
Concen: 18.675 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion:130 Resp: 57273
Ion Ratio Lower Upper
130 100
95 105.1 0.0 195.4









#45
1,2-Dichloropropane
Concen: 18.283 ug/l
RT: 9.623 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 63 Resp: 63647
Ion Ratio Lower Upper
63 100
65 31.1 24.1 36.1



#46
Dibromomethane
Concen: 17.235 ug/l
RT: 9.712 min Scan# 1319
Delta R.T. 0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 93 Resp: 41469
Ion Ratio Lower Upper
93 100
95 73.4 64.8 97.2
174 84.6 73.4 110.0









#47
Bromodichloromethane
Concen: 16.701 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion:	83	Resp:	84750
Ion	Ratio	Lower	Upper
83	100		
85	65.2	51.9	77.9
127	7.9	7.5	11.3



#48
Methyl methacrylate
Concen: 16.227 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion:	41	Resp:	69783
Ion	Ratio	Lower	Upper
41	100		
69	81.3	77.9	116.9
39	62.3	50.9	76.3





#49

1,4-Dioxane

Concen: 310.864 ug/l

RT: 9.700 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :

VN0426WBS02





Tgt Ion: 88 Resp: 20319

Ion Ratio Lower Upper

88 100

43 0.0 19.0 28.6#

58 78.3 50.6 76.0#



#50

Toluene-d8

Concen: 51.246 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN081872.D




Acq: 26 Apr 2024 12:47


Tgt Ion: 98 Resp: 548184

Ion Ratio Lower Upper

98 100

100 65.5 51.8 77.6





#51

4-Methyl-2-Pentanone

Concen: 84.116 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :





VN0426WBS02

Tgt Ion: 43 Resp: 426569

Ion Ratio Lower Upper

43 100

58 37.8 33.2 49.8



#52

Toluene

Concen: 18.809 ug/l

RT: 10.629 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VN081872.D




Acq: 26 Apr 2024 12:47


Tgt Ion: 92 Resp: 149832

Ion Ratio Lower Upper

92 100

91 169.3 136.7 205.1









#53
 t-1,3-Dichloropropene
 Concen: 17.268 ug/l
 RT: 10.835 min Scan# 1510
 Delta R.T. -0.000 min
 Lab File: VN081872.D
 Acq: 26 Apr 2024 12:47




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0426WBS02


Tgt Ion: 75 Resp: 95343
 Ion Ratio Lower Upper
 75 100
 77 30.4 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 18.367 ug/l
 RT: 10.312 min Scan# 1421
 Delta R.T. -0.000 min
 Lab File: VN081872.D
 Acq: 26 Apr 2024 12:47

Tgt Ion: 75 Resp: 104522
 Ion Ratio Lower Upper
 75 100
 77 30.5 25.0 37.4
 39 57.8 34.8 52.2#









#55
1,1,2-Trichloroethane
Concen: 17.409 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 97	Resp: 54541
Ion Ratio	Lower Upper
97 100	
83 88.0	70.4 105.6
85 53.5	45.6 68.4
99 59.0	50.9 76.3



#56
Ethyl methacrylate
Concen: 16.850 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 69	Resp: 100519
Ion Ratio	Lower Upper
69 100	
41 71.1	46.6 70.0#
39 43.6	26.5 39.7#









#57
1,3-Dichloropropane
Concen: 17.952 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 76 Resp: 99989
Ion Ratio Lower Upper
76 100
78 32.4 25.8 38.6



#58
2-Chloroethyl Vinyl ether
Concen: 82.892 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 63 Resp: 225270
Ion Ratio Lower Upper
63 100
106 26.5 24.6 36.8









#59
2-Hexanone
Concen: 85.746 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 43 Resp: 323606
Ion Ratio Lower Upper
43 100
58 52.8 28.3 84.9



#60
Dibromochloromethane
Concen: 17.741 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 129 Resp: 58869
Ion Ratio Lower Upper
129 100
127 77.5 38.3 114.8









#61
1,2-Dibromoethane
Concen: 17.862 ug/l
RT: 11.470 min Scan# 1618
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion:107 Resp: 53513
Ion Ratio Lower Upper
107 100
109 93.7 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 45.874 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 95 Resp: 193985
Ion Ratio Lower Upper
95 100
174 66.0 0.0 160.6
176 62.8 0.0 152.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :

VN0426WBS02





Tgt Ion:117 Resp: 377597

Ion Ratio Lower Upper

117 100

82 60.6 44.6 66.8

119 31.4 25.9 38.9



#64

Tetrachloroethene

Concen: 19.664 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Tgt Ion:164 Resp: 53927




Ion Ratio Lower Upper


164 100

166 129.0 97.6 146.4

129 92.4 73.7 110.5

131 100.5 75.8 113.6









#65
Chlorobenzene
Concen: 18.471 ug/l
RT: 11.894 min Scan# 1
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion:112 Resp: 152308
Ion Ratio Lower Upper
112 100
114 31.1 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 18.098 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion:131 Resp: 48500
Ion Ratio Lower Upper
131 100
133 101.7 47.5 142.6
119 71.2 32.9 98.7









#67
Ethyl Benzene
Concen: 18.328 ug/l
RT: 11.964 min Scan# 1702
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 91 Resp: 281068
Ion Ratio Lower Upper
91 100
106 30.0 25.3 37.9



#68
m/p-Xylenes
Concen: 37.840 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 106 Resp: 206985
Ion Ratio Lower Upper
106 100
91 213.8 159.4 239.0









#69
o-Xylene
Concen: 19.697 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion:106 Resp: 105189
Ion Ratio Lower Upper
106 100
91 215.6 104.7 314.1



#70
Styrene
Concen: 18.856 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion:104 Resp: 174199
Ion Ratio Lower Upper
104 100
78 53.3 38.6 58.0
103 51.5 43.8 65.6









#71
Bromoform
Concen: 16.174 ug/l
RT: 12.576 min Scan# 1806
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion:173 Resp: 34076
Ion Ratio Lower Upper
173 100
175 50.6 24.6 74.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion:152 Resp: 161162
Ion Ratio Lower Upper
152 100
115 65.6 30.4 91.3
150 164.2 0.0 353.2









#73
Isopropylbenzene
Concen: 18.091 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion:105 Resp: 259227
Ion Ratio Lower Upper
105 100
120 26.1 13.4 40.1



#74
N-ethyl acetate
Concen: 16.343 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 43 Resp: 130375
Ion Ratio Lower Upper
43 100
70 40.5 39.0 58.4
55 23.1 21.3 31.9
61 23.5 20.5 30.7





#75

1,1,2,2-Tetrachloroethane

Concen: 16.107 ug/l

RT: 12.941 min Scan# 1868

Delta R.T. 0.006 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :

VN0426WBS02





Tgt Ion: 83 Resp: 69306

Ion Ratio Lower Upper

83 100

131 9.7 5.5 16.4

85 66.1 32.9 98.6



#76

1,2,3-Trichloropropane

Concen: 24.775 ug/l

RT: 12.994 min Scan# 1877

Delta R.T. -0.000 min

Lab File: VN081872.D




Acq: 26 Apr 2024 12:47


Tgt Ion: 75 Resp: 103523

Ion Ratio Lower Upper

75 100

77 130.6 90.5 271.3





#77

Bromobenzene

Concen: 18.537 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :

VN0426WBS02





Tgt Ion:156 Resp: 55583

Ion Ratio Lower Upper

156 100

77 243.1 100.2 300.6

158 95.8 49.1 147.3



#78

n-propylbenzene

Concen: 17.825 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. -0.000 min

Lab File: VN081872.D




Acq: 26 Apr 2024 12:47


Tgt Ion: 91 Resp: 310249

Ion Ratio Lower Upper

91 100

120 22.7 11.7 35.0









#79
2-Chlorotoluene
Concen: 17.130 ug/l
RT: 13.123 min Scan# 1899
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 91 Resp: 189686
Ion Ratio Lower Upper
91 100
126 30.1 17.5 52.5



#80
1,3,5-Trimethylbenzene
Concen: 17.511 ug/l
RT: 13.170 min Scan# 1907
Delta R.T. -0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 105 Resp: 212321
Ion Ratio Lower Upper
105 100
120 47.5 24.8 74.3









#81
trans-1,4-Dichloro-2-butene
Concen: 17.406 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 75 Resp: 35494
Ion Ratio Lower Upper
75 100
53 101.7 72.9 109.3
89 39.3 39.1 58.7



#82
4-Chlorotoluene
Concen: 17.327 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 91 Resp: 186280
Ion Ratio Lower Upper
91 100
126 30.3 17.4 52.3









#83
tert-Butylbenzene
Concen: 17.961 ug/l
RT: 13.435 min Scan# 1952
Delta R.T. -0.006 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion:119 Resp: 181360
Ion Ratio Lower Upper
119 100
91 71.7 32.6 97.8
134 25.5 12.4 37.0



#84
1,2,4-Trimethylbenzene
Concen: 17.803 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion:105 Resp: 215569
Ion Ratio Lower Upper
105 100
120 43.3 22.9 68.7









#85
 sec-Butylbenzene
 Concen: 18.545 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. -0.000 min
 Lab File: VN081872.D
 Acq: 26 Apr 2024 12:47




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0426WBS02


Tgt Ion:105 Resp: 263768
 Ion Ratio Lower Upper
 105 100
 134 19.8 10.2 30.4



#86
 p-Isopropyltoluene
 Concen: 18.546 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. -0.000 min
 Lab File: VN081872.D
 Acq: 26 Apr 2024 12:47

Tgt Ion:119 Resp: 211808
 Ion Ratio Lower Upper
 119 100
 134 25.8 13.2 39.5
 91 26.1 11.8 35.3





#87

1,3-Dichlorobenzene

Concen: 17.952 ug/l

RT: 13.729 min Scan# 2003

Delta R.T. -0.006 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :

VN0426WBS02





Tgt Ion:146 Resp: 100436

Ion Ratio Lower Upper

146 100

111 44.5 20.8 62.5

148 64.2 32.1 96.3



#88

1,4-Dichlorobenzene

Concen: 18.294 ug/l

RT: 13.811 min Scan# 2016

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47




Tgt Ion:146 Resp: 100791


Ion Ratio Lower Upper

146 100

111 45.2 19.3 57.9

148 64.9 31.3 93.9









#89
n-Butylbenzene
Concen: 17.861 ug/l
RT: 14.058 min Scan# 2058
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47




Instrument :
MSVOA_N
ClientSampleId :
VN0426WBS02


Tgt Ion: 91 Resp: 189868
Ion Ratio Lower Upper
91 100
92 54.0 26.2 78.5
134 23.9 13.2 39.5



#90
Hexachloroethane
Concen: 17.913 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. -0.000 min
Lab File: VN081872.D
Acq: 26 Apr 2024 12:47

Tgt Ion: 117 Resp: 40243
Ion Ratio Lower Upper
117 100
201 61.8 33.3 99.8









#91
 1,2-Dichlorobenzene
 Concen: 18.498 ug/l
 RT: 14.106 min Scan# 2066
 Delta R.T. -0.000 min
 Lab File: VN081872.D
 Acq: 26 Apr 2024 12:47




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0426WBS02


Tgt Ion:146	Resp:	99211
Ion Ratio	Lower	Upper
146	100	
111	45.7	21.3 64.0
148	63.5	31.8 95.4



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 16.300 ug/l
 RT: 14.717 min Scan# 2170
 Delta R.T. -0.000 min
 Lab File: VN081872.D
 Acq: 26 Apr 2024 12:47

Tgt Ion: 75	Resp:	17076
Ion Ratio	Lower	Upper
75	100	
155	63.8	42.6 128.0
157	88.3	53.0 159.0









#93
 1,2,4-Trichlorobenzene
 Concen: 19.013 ug/l
 RT: 15.394 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: VN081872.D
 Acq: 26 Apr 2024 12:47




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0426WBS02


Tgt Ion:180	Resp:	57050
Ion Ratio	Lower	Upper
180	100	
182	95.8	46.7 140.0
145	30.9	16.2 48.6



#94
 Hexachlorobutadiene
 Concen: 19.586 ug/l
 RT: 15.505 min Scan# 2304
 Delta R.T. 0.006 min
 Lab File: VN081872.D
 Acq: 26 Apr 2024 12:47

Tgt Ion:225	Resp:	21661
Ion Ratio	Lower	Upper
225	100	
223	62.9	31.1 93.2
227	63.6	30.6 91.6





#95

Naphthalene

Concen: 17.687 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Instrument :

MSVOA_N

ClientSampleId :

VN0426WBS02





Tgt Ion:128 Resp: 200536

Ion Ratio Lower Upper

128 100

127 13.0 10.1 15.1

129 11.2 8.9 13.3



#96

1,2,3-Trichlorobenzene

Concen: 19.709 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. -0.000 min

Lab File: VN081872.D

Acq: 26 Apr 2024 12:47

Tgt Ion:180 Resp: 56779

Ion Ratio Lower Upper

180 100

182 95.4 46.0 137.8

145 35.1 17.1 51.3

