

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042624\
 Data File : VN081874.D
 Acq On : 26 Apr 2024 13:44
 Operator : JC\MD
 Sample : P2257-05DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41CDL

Quant Time: Apr 26 22:59:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	217879	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	407314	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	362377	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	146595	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	164239	44.664	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.320%
35) Dibromofluoromethane	8.165	113	120018	48.939	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	10.565	98	494310	51.322	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.640%
62) 4-Bromofluorobenzene	12.847	95	177198	46.540	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	93.080%
Target Compounds						
					Qvalue	
18) Methyl Acetate	5.030	43	10032	2.693	ug/l #	79
25) 2-Butanone	7.483	43	237976	114.022	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042624\
Data File : VN081874.D
Acq On : 26 Apr 2024 13:44
Operator : JC\MD
Sample : P2257-05DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41CDL

Quant Time: Apr 26 22:59:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

