

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042624\
 Data File : VN081882.D
 Acq On : 26 Apr 2024 16:56
 Operator : JC\MD
 Sample : P2282-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Apr 26 23:03:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	234145	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	433725	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	391107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	153778	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	173707	43.957	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.920%
35) Dibromofluoromethane	8.171	113	127112	48.676	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.360%
50) Toluene-d8	10.570	98	531372	51.810	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.620%
62) 4-Bromofluorobenzene	12.853	95	190586	47.009	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	94.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042624\
Data File : VN081882.D
Acq On : 26 Apr 2024 16:56
Operator : JC\MD
Sample : P2282-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Apr 26 23:03:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

