

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042624\
 Data File : VN081887.D
 Acq On : 26 Apr 2024 18:55
 Operator : JC\MD
 Sample : P2263-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 60156

Quant Time: Apr 26 23:05:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 26 05:26:49 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	247380	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	461895	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	400549	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	164959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	178217	42.685	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.380%
35) Dibromofluoromethane	8.165	113	133747	48.093	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.180%
50) Toluene-d8	10.565	98	555326	50.844	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.680%
62) 4-Bromofluorobenzene	12.847	95	201809	46.741	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	93.480%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.377	101	46816	15.896	ug/l	95
16) Acetone	4.424	43	68228	50.908	ug/l	97
27) cis-1,2-Dichloroethene	7.489	96	7797	2.042	ug/l	93
40) Benzene	8.612	78	53079	3.908	ug/l	97
44) Trichloroethene	9.353	130	95471	30.488	ug/l	91
51) 4-Methyl-2-Pentanone	10.453	43	6626	1.280	ug/l	90
52) Toluene	10.630	92	93106	11.447	ug/l	98
67) Ethyl Benzene	11.965	91	7524	0.463	ug/l #	81
68) m/p-Xylenes	12.071	106	6959	1.199	ug/l	99
69) o-Xylene	12.400	106	20664	3.648	ug/l	85
80) 1,3,5-Trimethylbenzene	13.176	105	4899	0.395	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042624\
Data File : VN081887.D
Acq On : 26 Apr 2024 18:55
Operator : JC\MD
Sample : P2263-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
60156

Quant Time: Apr 26 23:05:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042524W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 26 05:26:49 2024
Response via : Initial Calibration

