

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042719\
 Data File : VN055278.D
 Acq On : 26 Apr 2019 16:07
 Operator : JC/SP
 Sample : K2578-01 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WATER-COMP

Quant Time: Apr 27 03:29:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	367247	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	626721	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	554023	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142006	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	256195	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
35) Dibromofluoromethane	7.59	113	204916	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
50) Toluene-d8	10.09	98	762401	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
62) 4-Bromofluorobenzene	12.40	95	220332	44.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.22%	
Target Compounds						
3) Chloromethane	2.06	50	13760	2.375	ug/l	97
6) Chloroethane	2.71	64	12356	5.049	ug/l	90
16) Acetone	3.82	43	24877	15.139	ug/l	99
25) 2-Butanone	6.86	43	10250	4.832	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042719\
Data File : VN055278.D
Acq On : 26 Apr 2019 16:07
Operator : JC/SP
Sample : K2578-01 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WATER-COMP

Quant Time: Apr 27 03:29:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:52:53 2019
Response via : Initial Calibration

