

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042719\
 Data File : VN055267.D
 Acq On : 26 Apr 2019 11:27
 Operator : JC/SP
 Sample : VN0427WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0427WBL01

Quant Time: Apr 27 02:17:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:52:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	379609	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	655879	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	590735	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	149040	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	264229	51.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.86%	
35) Dibromofluoromethane	7.59	113	212384	47.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.34%	
50) Toluene-d8	10.09	98	810431	49.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.18%	
62) 4-Bromofluorobenzene	12.40	95	231141	44.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.42%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042719\
Data File : VN055267.D
Acq On : 26 Apr 2019 11:27
Operator : JC/SP
Sample : VN0427WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0427WBL01

Quant Time: Apr 27 02:17:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042219W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:52:53 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.00 min

Lab File: VN055267.D

Acq: 26 Apr 2019 11:27

Instrument :

MSVOA_N

ClientSampleId :


VN0427WBL01

Tot Ion:168 Resp: 379609

Ion Ratio Lower Upper




168 100

99 54.9 42.0 63.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0



#33

1,2-Dichloroethane-d4

Concen: 51.425 ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN055267.D


Acq: 26 Apr 2019 11:27

Tgt Ion: 65 Resp: 264229

Ion Ratio Lower Upper



65 100


67 53.5 0.0 108.6



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. -0.00 min

Lab File: VN055267.D

Acq: 26 Apr 2019 11:27

Instrument :

MSVOA_N

ClientSampleId :

VN0427WBL01


Tot Ion:114 Resp: 655879

Ion Ratio Lower Upper

114 100

63 23.3 0.0 45.2




88 16.2 0.0 32.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 47.666 ug/l

RT: 7.59 min Scan# 1810

Delta R.T. 0.00 min

Lab File: VN055267.D

Acq: 26 Apr 2019 11:27


Tgt Ion:113 Resp: 212384

Ion Ratio Lower Upper

113 100

111 104.1 81.0 121.4



192 21.3 16.2 24.4




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.090 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN055267.D

Acq: 26 Apr 2019 11:27

Instrument :

MSVOA_N

ClientSampleId :


VN0427WBL01

Tot Ion: 98 Resp: 810431

Ion Ratio Lower Upper

98 100

100 62.4 51.8 77.6



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN

10.09

400000



300000

200000

100000

0

Time-->



#62

4-Bromofluorobenzene

Concen: 44.212 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.00 min

Lab File: VN055267.D

Acq: 26 Apr 2019 11:27


Tgt Ion: 95 Resp: 231141

Ion Ratio Lower Upper

95 100

174 86.4 0.0 168.4

176 82.4 0.0 164.0



Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

12.40


150000


100000

50000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN055267.D
Acq: 26 Apr 2019 11:27

Instrument :
MSVOA_N
ClientSampleId :
VN0427WBL01

Tot Ion:117 Resp: 590735
Ion Ratio Lower Upper
117 100
82 56.7 45.7 68.5
119 31.5 25.4 38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3598
Delta R.T. 0.00 min
Lab File: VN055267.D
Acq: 26 Apr 2019 11:27

Tgt Ion:152 Resp: 149040
Ion Ratio Lower Upper
152 100
115 58.1 28.9 86.7
150 160.3 0.0 346.6

