

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042721\
 Data File : VN066794.D
 Acq On : 27 Apr 2021 20:25
 Operator : JC/MD
 Sample : M2170-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DN-1

Manual Integrations
 APPROVED

MMDadoda
 4/28/2021 2:38:55 PM

Quant Time: Apr 28 08:14:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	344230	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	528405	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.348	117	480259	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	211228	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	199944	45.21	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 90.42%			
35) Dibromofluoromethane	7.513	113	165704	50.80	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.60%			
50) Toluene-d8	10.029	98	650989	50.85	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 101.70%			
62) 4-Bromofluorobenzene	12.346	95	205389	47.41	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 94.82%			
Target Compounds						Qvalue
16) Acetone	3.758	43	34272	20.54	ug/l	93
18) Methyl Acetate	4.254	43	5339m	2.27	ug/l	
20) Methylene Chloride	4.474	84	22553	5.33	ug/l	97
43) Isopropyl Acetate	8.092	43	26017	3.14	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042721\
Data File : VN066794.D
Acq On : 27 Apr 2021 20:25
Operator : JC/MD
Sample : M2170-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DN-1

Manual Integrations
APPROVED

MMDadoda
4/28/2021 2:38:55 PM

Quant Time: Apr 28 08:14:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

