

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042721\
 Data File : VN066798.D
 Acq On : 27 Apr 2021 22:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 28 08:15:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	48.444	3.1	89	0.00
3 P	Chloromethane	50.000	50.175	-0.3	97	0.00
4 C	Vinyl Chloride	50.000	51.177	-2.4#	94	0.00
5 T	Bromomethane	50.000	47.195	5.6	89	0.00
6 T	Chloroethane	50.000	49.199	1.6	91	0.00
7 T	Trichlorofluoromethane	50.000	47.496	5.0	89	0.00
8 T	Diethyl Ether	50.000	50.115	-0.2	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.607	2.8	91	0.00
10 T	Methyl Iodide	50.000	51.075	-2.2	93	0.00
11 T	Tert butyl alcohol	250.000	243.757	2.5	91	0.00
12 CM	1,1-Dichloroethene	50.000	49.929	0.1#	92	0.00
13 T	Acrolein	250.000	224.819	10.1	84	0.00
14 T	Allyl chloride	50.000	47.572	4.9	87	0.00
15 T	Acrylonitrile	250.000	260.928	-4.4	89	0.00
16 T	Acetone	250.000	215.548	13.8	84	0.00
17 T	Carbon Disulfide	50.000	49.461	1.1	92	0.00
18 T	Methyl Acetate	50.000	46.594	6.8	88	0.00
19 T	Methyl tert-butyl Ether	50.000	49.168	1.7	89	0.00
20 T	Methylene Chloride	50.000	49.293	1.4	94	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.909	-1.8	92	0.00
22 T	Diisopropyl ether	50.000	48.907	2.2	88	0.00
23 T	Vinyl Acetate	250.000	244.150	2.3	85	0.00
24 P	1,1-Dichloroethane	50.000	50.233	-0.5	92	0.00
25 T	2-Butanone	250.000	247.449	1.0	87	0.00
26 T	2,2-Dichloropropane	50.000	42.797	14.4	80	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.812	-1.6	92	0.00
28 T	Bromochloromethane	50.000	48.832	2.3	92	0.00
29 T	Tetrahydrofuran	250.000	249.564	0.2	87	0.00
30 C	Chloroform	50.000	49.107	1.8#	90	0.00
31 T	Cyclohexane	50.000	46.584	6.8	89	0.00
32 T	1,1,1-Trichloroethane	50.000	48.621	2.8	89	0.00
33 S	1,2-Dichloroethane-d4	50.000	45.537	8.9	88	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	92	0.00
35 S	Dibromofluoromethane	50.000	51.314	-2.6	95	0.00
36 T	1,1-Dichloropropene	50.000	50.645	-1.3	89	0.00
37 T	Ethyl Acetate	50.000	47.473	5.1	84	0.00
38 T	Carbon Tetrachloride	50.000	50.233	-0.5	89	0.00
39 T	Methylcyclohexane	50.000	51.926	-3.9	90	0.00
40 TM	Benzene	50.000	52.450	-4.9	92	0.00
41 T	Methacrylonitrile	50.000	45.798	8.4	75	0.00
42 TM	1,2-Dichloroethane	50.000	48.591	2.8	86	0.00
43 T	Isopropyl Acetate	50.000	49.302	1.4	85	0.00
44 TM	Trichloroethene	50.000	53.410	-6.8	94	0.00
45 C	1,2-Dichloropropane	50.000	52.463	-4.9#	93	0.00
46 T	Dibromomethane	50.000	52.373	-4.7	91	0.00
47 T	Bromodichloromethane	50.000	51.666	-3.3	90	0.00
48 T	Methyl methacrylate	50.000	51.752	-3.5	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042721\
 Data File : VN066798.D
 Acq On : 27 Apr 2021 22:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 28 08:15:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1233.923	-23.4	100	0.00
50 S	Toluene-d8	50.000	50.300	-0.6	94	0.00
51 T	4-Methyl-2-Pentanone	250.000	263.016	-5.2	86	0.00
52 CM	Toluene	50.000	53.065	-6.1#	92	0.00
53 T	t-1,3-Dichloropropene	50.000	52.257	-4.5	88	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.411	-4.8	90	0.00
55 T	1,1,2-Trichloroethane	50.000	54.106	-8.2	93	0.00
56 T	Ethyl methacrylate	50.000	47.425	5.2	90	0.00
57 T	1,3-Dichloropropane	50.000	53.339	-6.7	92	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	261.186	-4.5	94	0.00
59 T	2-Hexanone	250.000	224.093	10.4	85	0.00
60 T	Dibromochloromethane	50.000	54.418	-8.8	92	0.00
61 T	1,2-Dibromoethane	50.000	54.754	-9.5	92	0.00
62 S	4-Bromofluorobenzene	50.000	52.582	-5.2	95	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	51.916	-3.8	96	0.00
65 PM	Chlorobenzene	50.000	51.510	-3.0	92	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.138	-4.3	93	0.00
67 C	Ethyl Benzene	50.000	52.818	-5.6#	91	0.00
68 T	m/p-Xylenes	100.000	109.145	-9.1	92	0.00
69 T	o-Xylene	50.000	53.185	-6.4	92	0.00
70 T	Styrene	50.000	47.541	4.9	92	0.00
71 P	Bromoform	50.000	56.580	-13.2	93	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	99	0.00
73 T	Isopropylbenzene	50.000	49.225	1.5	91	0.00
74 T	N-amyl acetate	50.000	49.380	1.2	81	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.121	5.8	89	0.00
76 T	1,2,3-Trichloropropane	50.000	49.542	0.9	88	0.00
77 T	Bromobenzene	50.000	50.693	-1.4	95	0.00
78 T	n-propylbenzene	50.000	50.697	-1.4	91	0.00
79 T	2-Chlorotoluene	50.000	48.418	3.2	93	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.616	0.8	93	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	41.562	16.9	81	0.00
82 T	4-Chlorotoluene	50.000	50.285	-0.6	93	0.00
83 T	tert-Butylbenzene	50.000	49.059	1.9	94	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.211	-2.4	93	0.00
85 T	sec-Butylbenzene	50.000	50.909	-1.8	93	0.00
86 T	p-Isopropyltoluene	50.000	51.894	-3.8	93	0.00
87 T	1,3-Dichlorobenzene	50.000	51.955	-3.9	96	0.00
88 T	1,4-Dichlorobenzene	50.000	49.888	0.2	100	0.00
89 T	n-Butylbenzene	50.000	48.939	2.1	89	0.00
90 T	Hexachloroethane	50.000	48.885	2.2	96	0.00
91 T	1,2-Dichlorobenzene	50.000	51.167	-2.3	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	45.920	8.2	93	0.00
93 T	1,2,4-Trichlorobenzene	50.000	54.106	-8.2	98	0.00
94 T	Hexachlorobutadiene	50.000	52.519	-5.0	104	0.00
95 T	Naphthalene	50.000	50.417	-0.8	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042721\
Data File : VN066798.D
Acq On : 27 Apr 2021 22:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 28 08:15:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.654	-5.3	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6