

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
 Data File : VN072218.D
 Acq On : 27 Apr 2022 16:05
 Operator : JC\MD
 Sample : N2575-06 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-95-98

Quant Time: Apr 28 06:04:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	282898	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	538215	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	569550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	200954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	218947	50.412	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.820%	
35) Dibromofluoromethane	8.022	113	163224	47.382	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 94.760%	
50) Toluene-d8	10.439	98	685566	50.647	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.300%	
62) 4-Bromofluorobenzene	12.727	95	257845	57.389	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 114.780%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	5.110	84	2017	0.562	ug/l	# 74
25) 2-Butanone	7.340	43	5727	1.920	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
Data File : VN072218.D
Acq On : 27 Apr 2022 16:05
Operator : JC\MD
Sample : N2575-06 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-95-98

Quant Time: Apr 28 06:04:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

