

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN042722\  
 Data File : VN072220.D  
 Acq On : 27 Apr 2022 16:53  
 Operator : JC\MD  
 Sample : N2575-08 50X  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 31796

Quant Time: Apr 28 06:05:01 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N042622W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 26 18:50:11 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	8.080	168	287937	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	542598	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	593548	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	212978	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	224668	50.824	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.640%
35) Dibromofluoromethane	8.016	113	166053	47.814	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.620%
50) Toluene-d8	10.439	98	704537	51.628	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.260%
62) 4-Bromofluorobenzene	12.727	95	272364	60.130	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	120.260%
Target Compounds						
						Qvalue
20) Methylene Chloride	5.081	84	2819	0.772	ug/l	# 74
25) 2-Butanone	7.345	43	5644	1.860	ug/l	# 89
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN042722\  
 Data File : VN072220.D  
 Acq On : 27 Apr 2022 16:53  
 Operator : JC\MD  
 Sample : N2575-08 50X  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 31796

Quant Time: Apr 28 06:05:01 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N042622W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 26 18:50:11 2022  
 Response via : Initial Calibration

