

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
 Data File : VN072223.D
 Acq On : 27 Apr 2022 18:05
 Operator : JC\MD
 Sample : N2575-01 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31809

Quant Time: Apr 28 06:05:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	273730	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	521247	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	570281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	208451	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	214700	51.090	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.180%	
35) Dibromofluoromethane	8.016	113	160810	48.201	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 96.400%	
50) Toluene-d8	10.439	98	679795	51.855	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.700%	
62) 4-Bromofluorobenzene	12.727	95	265530	61.023	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 122.040%	
Target Compounds						
						Qvalue
16) Acetone	4.293	43	11579	6.391	ug/l	95
20) Methylene Chloride	5.087	84	2550	0.734	ug/l #	70
25) 2-Butanone	7.334	43	3288	1.140	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
Data File : VN072223.D
Acq On : 27 Apr 2022 18:05
Operator : JC\MD
Sample : N2575-01 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31809

Quant Time: Apr 28 06:05:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

