

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
 Data File : VN072226.D
 Acq On : 27 Apr 2022 19:17
 Operator : JC\MD
 Sample : N2575-06 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-95-98

Quant Time: Apr 28 06:06:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	280406	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	520546	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	554259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	208538	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	215545	50.070	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 100.140%		
35) Dibromofluoromethane	8.016	113	160619	48.208	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 96.420%		
50) Toluene-d8	10.439	98	673658	51.456	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 102.920%		
62) 4-Bromofluorobenzene	12.727	95	269969	62.127	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 124.260%		
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	6213	3.348	ug/l	# 80
25) 2-Butanone	7.340	43	25474	8.618	ug/l	89
52) Toluene	10.504	92	2036	0.226	ug/l	96
95) Naphthalene	15.504	128	3291	4.578	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
Data File : VN072226.D
Acq On : 27 Apr 2022 19:17
Operator : JC\MD
Sample : N2575-06 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-95-98

Quant Time: Apr 28 06:06:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

