

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
 Data File : VN072227.D
 Acq On : 27 Apr 2022 19:41
 Operator : JC\MD
 Sample : N2575-07 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31795

Quant Time: Apr 28 06:06:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	263108	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	498286	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	527342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	182749	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	204952	50.740	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	=	101.480%	
35) Dibromofluoromethane	8.016	113	151158	47.395	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	=	94.800%	
50) Toluene-d8	10.439	98	637680	50.884	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	=	101.760%	
62) 4-Bromofluorobenzene	12.727	95	241317	58.014	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	=	116.020%	
Target Compounds						
						Qvalue
16) Acetone	4.287	43	5134	2.948	ug/l	90
20) Methylene Chloride	5.098	84	2080	0.623	ug/l #	81
25) 2-Butanone	7.334	43	26265	9.470	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
Data File : VN072227.D
Acq On : 27 Apr 2022 19:41
Operator : JC\MD
Sample : N2575-07 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31795

Quant Time: Apr 28 06:06:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

