

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
 Data File : VN072228.D
 Acq On : 27 Apr 2022 20:06
 Operator : JC\MD
 Sample : N2575-08 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31796

Quant Time: Apr 28 06:07:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	271228	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	503139	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	549944	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	205600	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	212133	50.945	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.900%
35) Dibromofluoromethane	8.016	113	155502	48.287	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.580%
50) Toluene-d8	10.439	98	652437	51.559	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.120%
62) 4-Bromofluorobenzene	12.727	95	258010	61.429	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	122.860%
Target Compounds						
						Qvalue
16) Acetone	4.298	43	5995	3.339	ug/l	# 82
25) 2-Butanone	7.333	43	26175	9.155	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
Data File : VN072228.D
Acq On : 27 Apr 2022 20:06
Operator : JC\MD
Sample : N2575-08 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31796

Quant Time: Apr 28 06:07:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

