

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
 Data File : VN072229.D
 Acq On : 27 Apr 2022 20:30
 Operator : JC\MD
 Sample : N2575-09 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31797

Quant Time: Apr 28 06:07:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	293763	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	538487	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	583239	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	214612	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	229289	50.841	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.680%	
35) Dibromofluoromethane	8.022	113	168185	48.797	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.600%	
50) Toluene-d8	10.439	98	694555	51.285	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.560%	
62) 4-Bromofluorobenzene	12.727	95	272456	60.610	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 121.220%	
Target Compounds						
						Qvalue
16) Acetone	4.293	43	4884	2.512	ug/l	89
25) 2-Butanone	7.334	43	24724	7.984	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
Data File : VN072229.D
Acq On : 27 Apr 2022 20:30
Operator : JC\MD
Sample : N2575-09 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31797

Quant Time: Apr 28 06:07:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

