

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
 Data File : VN072230.D
 Acq On : 27 Apr 2022 20:54
 Operator : JC\MD
 Sample : N2575-10 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31798

Quant Time: Apr 28 06:07:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 26 18:50:11 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	269183	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.963	114	506383	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	545005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	198907	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	215415	52.126	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.260%	
35) Dibromofluoromethane	8.022	113	153339	47.310	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 94.620%	
50) Toluene-d8	10.439	98	650554	51.081	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.160%	
62) 4-Bromofluorobenzene	12.727	95	257101	60.820	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 121.640%	
Target Compounds						
						Qvalue
16) Acetone	4.310	43	5841	3.278	ug/l	# 88
25) 2-Butanone	7.340	43	19249	6.784	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042722\
Data File : VN072230.D
Acq On : 27 Apr 2022 20:54
Operator : JC\MD
Sample : N2575-10 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31798

Quant Time: Apr 28 06:07:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042622W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 26 18:50:11 2022
Response via : Initial Calibration

