

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042723\
 Data File : VN077485.D
 Acq On : 27 Apr 2023 09:49
 Operator : JC\MD
 Sample : VN0427MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0427MBL01

Quant Time: Apr 28 03:21:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	582985	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1087769	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	978013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	339808	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	386586	50.818	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.640%	
35) Dibromofluoromethane	8.172	113	333838	51.062	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.120%	
50) Toluene-d8	10.566	98	1322785	52.790	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.580%	
62) 4-Bromofluorobenzene	12.848	95	427508	52.147	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042723\
Data File : VN077485.D
Acq On : 27 Apr 2023 09:49
Operator : JC\MD
Sample : VN0427MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0427MBL01

Quant Time: Apr 28 03:21:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

