

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042723\
 Data File : VN077490.D
 Acq On : 27 Apr 2023 11:59
 Operator : JC\MD
 Sample : 02520-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-042623

Quant Time: Apr 28 03:23:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	581715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1099623	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	969937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	337663	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	389098	51.260	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.520%
35) Dibromofluoromethane	8.172	113	336149	50.861	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.720%
50) Toluene-d8	10.572	98	1323014	52.230	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.460%
62) 4-Bromofluorobenzene	12.854	95	405293	48.904	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.800%
Target Compounds						
3) Chloromethane	2.378	50	1692	0.290	ug/l #	80
16) Acetone	4.449	43	13220	5.474	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042723\
Data File : VN077490.D
Acq On : 27 Apr 2023 11:59
Operator : JC\MD
Sample : 02520-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-01-042623

Quant Time: Apr 28 03:23:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

