

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042723\
 Data File : VN077491.D
 Acq On : 27 Apr 2023 12:24
 Operator : JC\MD
 Sample : 02520-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWOW-06B-40-45-042623

Quant Time: Apr 28 03:24:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	605739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1117764	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	994738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	365277	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	395414	50.026	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.060%	
35) Dibromofluoromethane	8.172	113	353100	52.558	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.120%	
50) Toluene-d8	10.572	98	1343456	52.177	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.360%	
62) 4-Bromofluorobenzene	12.854	95	424741	50.419	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.840%	
Target Compounds						
						Qvalue
3) Chloromethane	2.372	50	1905	0.314	ug/l #	82
4) Vinyl Chloride	2.525	62	3165	0.391	ug/l #	65
12) 1,1-Dichloroethene	4.354	96	256071	37.881	ug/l	95
21) trans-1,2-Dichloroethene	5.790	96	12670	1.733	ug/l #	71
24) 1,1-Dichloroethane	6.578	63	109703	8.985	ug/l #	96
27) cis-1,2-Dichloroethene	7.495	96	501048	57.878	ug/l	90
30) Chloroform	7.978	83	82539	5.998	ug/l	93
40) Benzene	8.613	78	10101	0.342	ug/l #	87
42) 1,2-Dichloroethane	8.678	62	42949	4.229	ug/l	96
44) Trichloroethene	9.354	130	57970	7.657	ug/l	100
47) Bromodichloromethane	9.895	83	6220	0.572	ug/l	97
49) 1,4-Dioxane	9.701	88	7383	50.678	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042723\
Data File : VN077491.D
Acq On : 27 Apr 2023 12:24
Operator : JC\MD
Sample : 02520-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWOW-06B-40-45-042623

Quant Time: Apr 28 03:24:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

