

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042723\
 Data File : VN077492.D
 Acq On : 27 Apr 2023 12:48
 Operator : JC\MD
 Sample : 02520-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWOW-06B-40-45-042623-FD

Quant Time: Apr 28 03:24:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:50:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	580709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1085920	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	960906	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	350208	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	388373	51.253	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.500%	
35) Dibromofluoromethane	8.172	113	337609	51.726	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.460%	
50) Toluene-d8	10.572	98	1287859	51.484	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.960%	
62) 4-Bromofluorobenzene	12.854	95	416378	50.876	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.760%	
Target Compounds						Qvalue
3) Chloromethane	2.366	50	1313	0.226	ug/l	92
4) Vinyl Chloride	2.531	62	4226	0.545	ug/l	92
12) 1,1-Dichloroethene	4.348	96	251992	38.885	ug/l	98
16) Acetone	4.460	43	9409	3.903	ug/l #	69
21) trans-1,2-Dichloroethene	5.795	96	12633	1.803	ug/l	89
24) 1,1-Dichloroethane	6.584	63	107057	9.146	ug/l	97
27) cis-1,2-Dichloroethene	7.495	96	510753	61.542	ug/l	90
30) Chloroform	7.972	83	80036	6.067	ug/l	95
40) Benzene	8.607	78	9822	0.342	ug/l #	89
42) 1,2-Dichloroethane	8.678	62	43544	4.413	ug/l	96
44) Trichloroethene	9.360	130	58889	8.007	ug/l	92
47) Bromodichloromethane	9.895	83	5493	0.520	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042723\
Data File : VN077492.D
Acq On : 27 Apr 2023 12:48
Operator : JC\MD
Sample : 02520-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWOW-06B-40-45-042623-FD

Quant Time: Apr 28 03:24:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042423W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:50:59 2023
Response via : Initial Calibration

