

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042820\
 Data File : VN061163.D
 Acq On : 28 Apr 2020 14:54
 Operator : JC/MD
 Sample : L2386-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB180-GW-558-560

Quant Time: Apr 28 20:07:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 28 02:20:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	106459	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	190547	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	183588	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	82249	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	85518	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
35) Dibromofluoromethane	7.55	113	57492	51.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.04%	
50) Toluene-d8	10.06	98	235501	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
62) 4-Bromofluorobenzene	12.38	95	89674	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
Target Compounds						
16) Acetone	3.78	43	6464	4.497	ug/l	93
44) Trichloroethene	8.80	130	3304	2.151	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042820\
Data File : VN061163.D
Acq On : 28 Apr 2020 14:54
Operator : JC/MD
Sample : L2386-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB180-GW-558-560

Quant Time: Apr 28 20:07:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 28 02:20:26 2020
Response via : Initial Calibration

