

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042820\
 Data File : VN061165.D
 Acq On : 28 Apr 2020 15:41
 Operator : JC/MD
 Sample : L2386-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB180-GW-538-540

Quant Time: Apr 28 20:10:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 28 02:20:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	102234	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	183285	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	175229	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	78878	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	81662	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
35) Dibromofluoromethane	7.55	113	56313	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
50) Toluene-d8	10.06	98	226048	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	12.37	95	85171	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
Target Compounds						
16) Acetone	3.78	43	6759	5.398	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042820\
Data File : VN061165.D
Acq On : 28 Apr 2020 15:41
Operator : JC/MD
Sample : L2386-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB180-GW-538-540

Quant Time: Apr 28 20:10:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 28 02:20:26 2020
Response via : Initial Calibration

