

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042820\
 Data File : VN061166.D
 Acq On : 28 Apr 2020 16:04
 Operator : JC/MD
 Sample : L2412-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB180-GW-588-590

Quant Time: Apr 29 08:34:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 28 02:20:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	100126	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	179764	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	174096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	78524	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	80984	51.12	ug/l	0.00
Spiked Amount			Recovery	=	102.24%	
35) Dibromofluoromethane	7.55	113	54473	51.25	ug/l	0.00
Spiked Amount			Recovery	=	102.50%	
50) Toluene-d8	10.06	98	224215	51.74	ug/l	0.00
Spiked Amount			Recovery	=	103.48%	
62) 4-Bromofluorobenzene	12.38	95	84898	52.06	ug/l	0.00
Spiked Amount			Recovery	=	104.12%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	3.72	101	6426	6.053	ug/l	98
12) 1,1-Dichloroethene	3.70	96	2932	2.775	ug/l	97
16) Acetone	3.78	43	6136	4.591	ug/l	94
27) cis-1,2-Dichloroethene	6.79	96	3977	2.946	ug/l	87
30) Chloroform	7.33	83	6528	2.714	ug/l	96
44) Trichloroethene	8.80	130	613190	423.096	ug/l	99
55) 1,1,2-Trichloroethane	10.54	97	2764	2.146	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042820\
Data File : VN061166.D
Acq On : 28 Apr 2020 16:04
Operator : JC/MD
Sample : L2412-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB180-GW-588-590

Quant Time: Apr 29 08:34:46 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 28 02:20:26 2020
Response via : Initial Calibration

