

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN042820\
 Data File : VN061167.D
 Acq On : 28 Apr 2020 16:27
 Operator : JC/MD
 Sample : L2412-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB180-GW-618-620

Quant Time: Apr 29 08:39:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 28 02:20:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	102158	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	182748	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	176007	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	80089	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	81477	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
35) Dibromofluoromethane	7.55	113	55416	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
50) Toluene-d8	10.06	98	225768	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
62) 4-Bromofluorobenzene	12.38	95	85587	51.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.26%	
Target Compounds						
16) Acetone	3.78	43	6110	4.345	ug/l	96
44) Trichloroethene	8.80	130	1934	1.313	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN042820\
Data File : VN061167.D
Acq On : 28 Apr 2020 16:27
Operator : JC/MD
Sample : L2412-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB180-GW-618-620

Quant Time: Apr 29 08:39:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 28 02:20:26 2020
Response via : Initial Calibration

