


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN042820\
 Data File : VN061157.D
 Acq On : 28 Apr 2020 12:16
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N042720W.M
 Title : SW846 8260
 Last Update : Wed Apr 29 08:14:26 2020



AutoFind: Scans 3353, 3354, 3355; Background Corrected with Scan 3341

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.7	6564	PASS
75	95	30	60	56.9	15775	PASS
95	95	100	100	100.0	27728	PASS
96	95	5	9	7.1	1960	PASS
173	174	0.00	2	0.3	67	PASS
174	95	50	100	71.2	19746	PASS
175	174	5	9	7.5	1486	PASS
176	174	95	101	97.2	19197	PASS
177	176	5	9	6.9	1326	PASS