

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066809.D
 Acq On : 28 Apr 2021 13:59
 Operator : JC/MD
 Sample : M2136-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 31075-76

Manual Integrations
 APPROVED

MMDadoda
 4/29/2021 5:52:11 PM

Quant Time: Apr 29 04:29:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	343547	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	514737	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	449573	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	189719	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.953	65	186860	42.33	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	84.66%		
35) Dibromofluoromethane	7.513	113	159292	50.13	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	100.26%		
50) Toluene-d8	10.029	98	615570	49.36	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.72%		
62) 4-Bromofluorobenzene	12.349	95	191358	45.35	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	90.70%		
Target Compounds						Qvalue
11) Tert butyl alcohol	4.702	59	21216	31.17	ug/l	# 80
16) Acetone	3.758	43	17119	10.28	ug/l	100
17) Carbon Disulfide	3.986	76	4749m	0.45	ug/l	
25) 2-Butanone	6.760	43	19327	8.26	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066809.D
Acq On : 28 Apr 2021 13:59
Operator : JC/MD
Sample : M2136-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
31075-76

Manual Integrations
APPROVED

MMDadoda
4/29/2021 5:52:11 PM

Quant Time: Apr 29 04:29:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

