

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066811.D
 Acq On : 28 Apr 2021 14:44
 Operator : JC/MD
 Sample : M2138-17
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-DDC-4-PS

Quant Time: Apr 29 04:29:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.594	168	354231	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.522	114	535573	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	460719	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	187725	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	191401	42.05	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	84.10%		
35) Dibromofluoromethane	7.516	113	165469	50.05	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	100.10%		
50) Toluene-d8	10.032	98	642709	49.53	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	99.06%		
62) 4-Bromofluorobenzene	12.349	95	169398	38.58	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	77.16%		
Target Compounds						
30) Chloroform	7.293	83	2804m	0.39	ug/l	Qvalue
64) Tetrachloroethene	10.573	164	12128	3.98	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066811.D
Acq On : 28 Apr 2021 14:44
Operator : JC/MD
Sample : M2138-17
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-DDC-4-PS

Quant Time: Apr 29 04:29:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

