

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066815.D
 Acq On : 28 Apr 2021 16:17
 Operator : JC/MD
 Sample : M2138-21
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 152140-14D

Quant Time: Apr 29 04:30:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	333372	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	510408	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	444855	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.293	152	193109	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	182648	42.64	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	85.28%
35) Dibromofluoromethane	7.513	113	158236	50.22	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.44%
50) Toluene-d8	10.032	98	606224	49.02	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.04%
62) 4-Bromofluorobenzene	12.349	95	180629	43.17	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.34%
Target Compounds						
64) Tetrachloroethene	10.579	164	1333	0.45	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066815.D
Acq On : 28 Apr 2021 16:17
Operator : JC/MD
Sample : M2138-21
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
152140-14D

Quant Time: Apr 29 04:30:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

