

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
 Data File : VN066819.D
 Acq On : 28 Apr 2021 17:51
 Operator : JC/MD
 Sample : PB136019ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB136019ZHE#01

Quant Time: Apr 29 04:31:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 27 16:02:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.591	168	320772	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.519	114	485075	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.351	117	454801	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.290	152	192784	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.956	65	174771	42.41	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	84.82%
35) Dibromofluoromethane	7.513	113	153485	51.26	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.52%
50) Toluene-d8	10.029	98	602857	51.29	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.58%
62) 4-Bromofluorobenzene	12.349	95	195096	49.06	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.12%
Target Compounds						
					Qvalue	
16) Acetone	3.756	43	22294	14.34	ug/l	97
20) Methylene Chloride	4.475	84	17325	4.40	ug/l	95
43) Isopropyl Acetate	8.095	43	28199	3.71	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN042821\
Data File : VN066819.D
Acq On : 28 Apr 2021 17:51
Operator : JC/MD
Sample : PB136019ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB136019ZHE#01

Quant Time: Apr 29 04:31:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N042721W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 27 16:02:50 2021
Response via : Initial Calibration

